

Efficacy Bulletin

DENTIRO® Wipes

PRODUCT DESCRIPTION

DENTIRO® Wipes are ready-to-use disinfectant wipes for the residue-free disinfection and cleaning of small surfaces of non-invasive medical devices such as operating tables, gurneys, IV poles, and dental chairs. The highly saturated wipes have a low alcohol content of less than 50%, which not only reduces the potential for allergies but also increases material compatibility. DENTIRO® Wipes are aldehyde-free and biodegradable.

INTRODUCTION

The product has been tested for compatibility with a variety of materials and devices, which are expected to come in contact with the product during its intended use. Testing was performed according to the below mentioned methods.

RELEVANT PHYSICAL AND CHEMICAL PROPERTIES

Composition of the moisturizing solution in 100 g:	31.3 g ethanol, 16.7 g 2-propanol
Physical state:	Liquid absorbed on fleece
pH-value:	Neutral
pH-value in aqueous solution:	Not applicable (ready-to-use solution)

Bacteria

ACETOBACTER ACETI - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		



ACHROMOBACTER XYLOSOXIDANS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.38
Test date:	07.11.2014		

BACILLUS LICHENIFORMIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.54
Test date:	27.02.2015		

BACILLUS PUMILIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

Dirty condition

Concentration: Undiluted. Time: 60 seconds
 Required log reduction: 5.00 Achieved log reduction: 5.00
 Test date: 03.02.2015

BACILLUS SUBTILIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

Dirty condition

Concentration: Undiluted. Time: 60 seconds
 Required log reduction: 5.00 Achieved log reduction: 5.00
 Test date: 03.02.2015

BACTEROIDES OVATUS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

Dirty condition

Concentration: Undiluted. Time: 60 seconds
 Required log reduction: 5.00 Achieved log reduction: 5.00
 Test date: 03.02.2015



BORDETELLA BRONCHISEPTICA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	7.15
Test date:	07.11.2014		

BORDETELLA PERTUSSIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.64
Test date:	07.11.2014		

BREVUNDIMONAS DIMINUTA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	7.23
Test date:	27.02.2015		

BURKHOLDERIA CEPACIA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.71
Test date:	07.11.2014		

CAMPYLOBACTER JEJUNI - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.08
Test date:	07.11.2014		



CHROMOBACTERIUM VIOLACEUM - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

CHRYSEOBACTERIUM INDOLOGENES - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

CITROBACTER FREUNDII - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

Dirty condition

Concentration: Undiluted. Time: 30 seconds
 Required log reduction: 5.00 Achieved log reduction: 7.48
 Test date: 07.11.2014

CLOSTRIDIUM DIFFICILE (VEGETATIVE) - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

Dirty condition

Concentration: Undiluted. Time: 30 seconds
 Required log reduction: 5.00 Achieved log reduction: 6.20
 Test date: 07.11.2014

CLOSTRIDIUM PERFRINGENS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

Dirty condition

Concentration: Undiluted. Time: 30 seconds
 Required log reduction: 5.00 Achieved log reduction: 5.96
 Test date: 27.02.2015



CORYNIBACTERIUM UREALYTICUM - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

ENTEROBACTER AEROGENES - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	7.00
Test date:	07.11.2014		

ENTEROBACTER CLOACAE - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.72
Test date:	07.11.2014		

ENTEROBACTER GERGOVIAE - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.90
Test date:	27.02.2015		

ENTEROCOCCUS CASSELI FLAVUS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.71
Test date:	27.02.2015		

ENTEROCOCCUS FAECALIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

30 seconds

Required log reduction: 5.00

Achieved log reduction:

6.72

Test date: 07.11.2014

ENTEROCOCCUS FAECIUM - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

60 seconds

Required log reduction: 5.00

Achieved log reduction:

5.00

Test date: 03.02.2015

ENTEROCOCCUS HIRAE - EN 1656

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in veterinary field-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 10 °C

Clean condition

Concentration: Undiluted.

Time:

60 seconds

ENTEROCOCCUS HIRAE - EN 13697

Chemical disinfectants and antiseptics-Quantitative non-porous surface test for the evaluation of bactericidal and/or fungicidal activity of chemical disinfectants used in food and industrial, domestic and institutional areas – Test method and requirements without mechanical action(phase 2,step 2)

Test method: Carrier test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

60 seconds

Required log reduction: 4.00

Achieved log reduction:

4.04

Test date: 15.04.2009

ENTEROCOCCUS HIRAE - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

30 seconds

Required log reduction: 5.00

Achieved log reduction:

5.14

Test date: 23.03.2009

ENTEROCOCCUS HIRAE - EN 14561

Chemical disinfectants and antiseptics-Quantitative carrier test for the evaluation of bactericidal activity of chemical disinfectants for instruments used in medical area – Test method requirements (phase 2, step 2)

Test method: Carrier test

Test temperature: 20 °C

Clean condition

Concentration: Undiluted.

Time:

60 seconds



ESCHERICHIA COLI O157:H7 - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.84
Test date:	07.11.2014		

FUSOBACTERIUM NUCLEATUM - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

GARDNERELLA VAGINALIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

HAEMOPHILUS INFLUENZA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	5.36
Test date:	07.11.2014		

KLEBSIELLA OXYTOCA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.79
Test date:	07.11.2014		



KLEBSIELLA PNEUMONIAE - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.51
Test date:	07.11.2014		

KYTOCOCCUS SEDENTARIUS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

LEGIONELLA PNEUMOPHILA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	5.18
Test date:	07.11.2014		

LISTERIA INNOCUA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.76
Test date:	27.02.2015		

LISTERIA MONOCYTOGENES - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.54
Test date:	07.11.2014		



MALESSEZIA FURFUR - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

MICROCOCCUS LUTEUS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

MORAXELLA CATARRHALIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.38
Test date:	07.11.2014		

NEISSERIA FLAVESCENS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

NEISSERIA MENINGITIDIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	5.45
Test date:	07.11.2014		

PORPHYSONOMAS GINGIVALIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

60 seconds

Required log reduction: 5.00

Achieved log reduction:

5.00

Test date: 03.02.2015

PREVOTELLA INTERMEDIA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

60 seconds

Required log reduction: 5.00

Achieved log reduction:

5.00

Test date: 03.02.2015

PROPIONIBACTERIUM ACNES - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition



Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

PROTEUS MIRABILIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	7.20
Test date:	07.11.2014		

PROTEUS VULGARIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.89
Test date:	07.11.2014		

PROTEUS VULGARIS - EN 14349

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in veterinary field on non-porous surfaces without mechanical action-Test method and requirements (phase 2, step 2)

Test method: Carrier test

Test temperature: 10 °C

Clean condition

Concentration: Undiluted. Time: 60 seconds

Required log reduction: 5.00 Achieved log reduction: 5.13

Test date: 20.10.2014

Dirty condition

Concentration: Time:

Required log reduction: Achieved log reduction:

Test date:

PROTEUS VULGARIS - EN 1656

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in veterinary field-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 10 °C

Clean condition

Concentration: Undiluted. Time: 60 seconds

Required log reduction: 5.00 Achieved log reduction: 5.55

Test date: 20.10.2014

Dirty condition

Concentration: Time:

Required log reduction: Achieved log reduction:

Test date:

PSEUDOMONAS AERUGINOSA - EN 1656

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in veterinary field-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 10 °C

Clean condition

Concentration: Undiluted. Time: 60 seconds

Required log reduction: 5.00 Achieved log reduction: 5.38



Test date: 20.10.2014

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

PSEUDOMONAS AERUGINOSA - EN 14349

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in veterinary field on non-porous surfaces without mechanical action-Test method and requirements (phase 2, step 2)

Test method: Carrier test
 Test temperature: 10 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.46
Test date:	20.10.2014		

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

PSEUDOMONAS AERUGINOSA - EN 13697

Chemical disinfectants and antiseptics-Quantitative non-porous surface test for the evaluation of bactericidal and/or fungicidal activity of chemical disinfectants used in food and industrial, domestic and institutional areas – Test method and requirements without mechanical action(phase 2,step 2)

Test method: Carrier test
 Test temperature: 20 °C

Clean condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	5.33
Test date:	15.04.2009		



PSEUDOMONAS AERUGINOSA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	5.15
Test date:	23.03.2009		

PSEUDOMONAS AERUGINOSA - EN 14561

Chemical disinfectants and antiseptics-Quantitative carrier test for the evaluation of bactericidal activity of chemical disinfectants for instruments used in medical area – Test method requirements (phase 2, step 2)

Test method: Carrier test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.73
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

PSEUDOMONAS AERUGINOSA - EN 13727

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity in the medical area – Test method requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.55



Test date: 20.10.2014

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

PSEUDOMONAS AERUGINOSA - EN 1040

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of basic bactericidal activity of chemical disinfectants and antiseptics-Test method and requirements (phase1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	5.03
Test date:	24.02.2009		

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

PSEUDOMONAS FLUORESCENS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		



RALSTONIA PICKETTII - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

SALMONELLA BONGORI - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.91
Test date:	07.11.2014		

SALMONELLA ENTERICA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition



Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	7.85
Test date:	07.11.2014		

SALMONELLA SERINFONTIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	7.04
Test date:	27.02.2015		

SALMONELLA TYPHIMURIUM - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.04
Test date:	07.11.2014		

SARCINA LUTEA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 60 seconds

Required log reduction: 5.00

Achieved log reduction: 5.00

Test date: 03.02.2015

SERRATIA MARCESCENS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 7.04

Test date: 07.11.2014

SPHINGOMONAS PAUCIMOBILIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition



Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

STAPHYLOCOCCUS AUREUS - EN 14349

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in veterinary field on non-porous surfaces without mechanical action-Test method and requirements (phase 2, step 2)

Test method: Carrier test
 Test temperature: 10 °C

Clean condition

Concentration:	Undiluted.	Time:	30 minutes
Required log reduction:	5.00	Achieved log reduction:	5.57
Test date:	20.10.2014		

Dirty condition

Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:

STAPHYLOCOCCUS AUREUS - EN 1656

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in veterinary field-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 10 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.40
Test date:	20.10.2014		

Dirty condition

Concentration: Time:
 Required log reduction: Achieved log reduction:
 Test date:



STAPHYLOCOCCUS AUREUS - EN 1040

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of basic bactericidal activity of chemical disinfectants and antiseptics-Test method and requirements (phase1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	5.21
Test date:	24.02.2009		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

STAPHYLOCOCCUS AUREUS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	5.07
Test date:	23.03.2009		

STAPHYLOCOCCUS AUREUS - EN 13727

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity in the medical area – Test method requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.51



Test date: 20.10.2014

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

STAPHYLOCOCCUS AUREUS - EN 14561

Chemical disinfectants and antiseptics-Quantitative carrier test for the evaluation of bactericidal activity of chemical disinfectants for instruments used in medical area – Test method requirements (phase 2, step 2)

Test method: Carrier test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.19
Test date:	20.10.2014		

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

STAPHYLOCOCCUS AUREUS - EN 13697

Chemical disinfectants and antiseptics-Quantitative non-porous surface test for the evaluation of bactericidal and/or fungicidal activity of chemical disinfectants used in food and industrial, domestic and institutional areas – Test method and requirements without mechanical action(phase 2,step 2)

Test method: Carrier test
 Test temperature: 20 °C

Clean condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	6.49
Test date:	15.04.2009		



STAPHYLOCOCCUS CAPITIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 5.88

Test date: 27.02.2015

STAPHYLOCOCCUS EPIDERMIDIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 6.67

Test date: 07.11.2014

STAPHYLOCOCCUS INTERMEDIUS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition



Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 7.00

Test date: 27.02.2015

STAPHYLOCOCCUS WARNERI - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 6.46

Test date: 27.02.2015

STENOTROPHOMONAS MALTOPHILA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 5.48

Test date: 27.02.2015

STREPTOCOCCUS AGALACTIAE - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

30 seconds

Required log reduction: 5.00

Achieved log reduction:

6.43

Test date: 07.11.2014

STREPTOCOCCUS PNEUMONIAE - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time:

30 seconds

Required log reduction: 5.00

Achieved log reduction:

5.65

Test date: 07.11.2014

STREPTOCOCCUS PYOGENES - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition



Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.20
Test date:	07.11.2014		

VIBRIO PARAHAEMOLYTICUS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.64
Test date:	07.11.2014		

YERSINIA ENTEROCOLITICA - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.57
Test date:	27.02.2015		



Bacteriah

ACINETOBACTER BAUMANNII - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 6.93

Test date: 07.11.2014

EXTENDED-SPECTRUM BETA-LACTAMASE ESCHERICHIA COLI - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 5.00

Achieved log reduction: 6.59

Test date: 07.11.2014

KOCURIA RHIZOPHILA STREPTOMYCIN RESISTANT - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C



Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

METHICILLIN-RESISTANT STAPHYLOCOCCUS AUREUS CLINICAL ISOLATE 4628-2 - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	5.00	Achieved log reduction:	6.75
Test date:	25.03.2009		

VANCOMYCIN-RESISTANT ENTEROCOCCUS FAECALIS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
----------------	------------	-------	------------



Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	30 minutes
Required log reduction:	4.00	Achieved log reduction:	4.31
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

ASPERGILLUS BRASILIENSIS - EN 1657

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of fungicidal or yeasticidal activity of chemical disinfectants and antiseptics used in veterinary area-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	30 minutes
Required log reduction:	4.00	Achieved log reduction:	4.50
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

ASPERGILLUS BRASILIENSIS - EN 13624

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of fungicidal activity of chemical disinfectants for instruments used in medical area – Test method requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	5 minutes
Required log reduction:	4.00	Achieved log reduction:	4.25
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	



Test date:

CANDIDA ALBICANS - EN 1275

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of basic fungicidal or basic yeasticidal activity of chemical disinfectants and antiseptics- Test method and requirements (phase 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	5 minutes
Required log reduction:	4.00	Achieved log reduction:	4.51
Test date:	24.02.2009		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

CANDIDA ALBICANS - EN 1650

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of fungicidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	4.04
Test date:	25.03.2009		

CANDIDA ALBICANS - EN 13697

Chemical disinfectants and antiseptics-Quantitative non-porous surface test for the evaluation of bactericidal and/or fungicidal activity of chemical disinfectants used in food and industrial, domestic and institutional areas – Test method and requirements without mechanical action(phase 2,step 2)

Test method: Carrier test
 Test temperature: 20 °C

Clean condition



Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	3.00	Achieved log reduction:	4.17
Test date:	20.10.2014		

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	3.00	Achieved log reduction:	5.28
Test date:	15.04.2009		

CANDIDA ALBICANS - EN 1657

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of fungicidal or yeasticidal activity of chemical disinfectants and antiseptics used in veterinary area-Test method and requirements (phase 2, step 1)

Test method:	Suspension test
Test temperature:	10 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	4.52
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

CANDIDA ALBICANS - EN 14562

Chemical disinfectants and antiseptics-Quantitative carrier test for the evaluation of fungicidal or yeasticidal activity for instruments used in medical area – Test method requirements (phase 2, step 2)

Test method:	Carrier test
Test temperature:	20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	4.31
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			



CANDIDA ALBICANS - EN 13624

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of fungicidal activity of chemical disinfectants for instruments used in medical area – Test method requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	4.52
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

CANDIDA GLABRATA - EN 1275

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of basic fungicidal or basic yeasticidal activity of chemical disinfectants and antiseptics- Test method and requirements (phase 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	6.04
Test date:	27.02.2015		

CANDIDA PARAPSILOSIS - EN 1275

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of basic fungicidal or basic yeasticidal activity of chemical disinfectants and antiseptics- Test method and requirements (phase 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	

Test date:

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	4.00
Test date:	03.02.2015		

CANDIDA UTILIS - EN 1275

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of basic fungicidal or basic yeasticidal activity of chemical disinfectants and antiseptics- Test method and requirements (phase 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	5.41
Test date:	27.02.2015		

SACCHAROMYCES CEREVISIAE - EN 1275

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of basic fungicidal or basic yeasticidal activity of chemical disinfectants and antiseptics- Test method and requirements (phase 1)

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	5.98
Test date:	27.02.2015		

TRICHODERMA VIRENS - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)



Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	5.11
Test date:	09.02.2015		

ZYGOSACCHAROMYCES ROUXII - EN 1276

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of bactericidal activity of chemical disinfectants and antiseptics used in food, industrial, domestic and institutional areas-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	5.00	Achieved log reduction:	5.00
Test date:	03.02.2015		

Mycobacteria

MYCOBACTERIUM AVIUM - EN 14348

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants and antiseptics in the medical area including instrument disinfectants-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	5.03



Test date: 15.05.2014

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

MYCOBACTERIUM AVIUM - EN 14204

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants and antiseptics used in veterinary area-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 10 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	4.97
Test date:	20.10.2014		

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

MYCOBACTERIUM AVIUM - EN 14563

Chemical disinfectants and antiseptics-Quantitative carrier test for the evaluation of mycobactericidal or tuberculocidal activity of chemicals used for instruments in medical area – Test method requirements (phase 2, step 2)

Test method: Carrier test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	5.78
Test date:	20.10.2014		

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

MYCOBACTERIUM BOVIS - EN 14348

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants and antiseptics in the medical area including instrument disinfectants-Test method and requirements (phase 2, step 1)



Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	4.00
Test date:	03.02.2015		

MYCOBACTERIUM CHELONAE - EN 14348

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants and antiseptics in the medical area including instrument disinfectants-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	6.28
Test date:	03.03.2014		

MYCOBACTERIUM SMEGMATIS - EN 14348

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants and antiseptics in the medical area including instrument disinfectants-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			



Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	6.41
Test date:	03.03.2015		

MYCOBACTERIUM TERRAE - EN 14348

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of mycobactericidal activity of chemical disinfectants and antiseptics in the medical area including instrument disinfectants-Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	6.52
Test date:	24.03.2009		

MYCOBACTERIUM TERRAE - EN 14563

Chemical disinfectants and antiseptics-Quantitative carrier test for the evaluation of mycobactericidal or tuberculocidal activity of chemicals used for instruments in medical area – Test method requirements (phase 2, step 2)

Test method: Carrier test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	5.61
Test date:	20.10.2014		

Dirty condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Viruses

ADENOVIRUS TYPE 5 - EN 14476

Chemical disinfectants and antiseptics- Viricidal quantitative suspension test for the evaluation chemical disinfectants and antiseptics used in human medicine- Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 60 seconds

Required log reduction: 4.00

Achieved log reduction: 4.00

Test date: 29.01.2014

BOVINE ENTEROVIRUS TYPE 1 - EN 14675

Chemical disinfectants and antiseptics-Quantitative suspension test for the evaluation of virucidal activity of chemical disinfectants and antiseptics used in veterinary area-Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 10 °C

Clean condition

Concentration: Undiluted.

Time: 60 seconds

Required log reduction: 4.00

Achieved log reduction: 5.00

Test date: 20.10.2014

Dirty condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

LACTOCOCCUS LACTIS SUBSP. LACTIS BACTERIOPHAGE P008 - EN 13610

Chemical disinfectants -Quantitative suspension test for the evaluation of virucidal activity against bacteriophages of chemical disinfectants used in food and industrial areas – Test method and requirements (phase 2, step 2)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration: Undiluted.

Time: 15 minutes

Required log reduction: 4.00

Achieved log reduction: 4.75



Test date: 20.10.2014

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

NOROVIRUS FCV - EN 14476

Chemical disinfectants and antiseptics- Viricidal quantitative suspension test for the evaluation chemical disinfectants and antiseptics used in human medicine- Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	5.00
Test date:	20.10.2014		

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

POLIOVIRUS TYPE 1 LSC-2AB - EN 14476

Chemical disinfectants and antiseptics- Viricidal quantitative suspension test for the evaluation chemical disinfectants and antiseptics used in human medicine- Test method and requirements (phase 2, step 1)

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	60 seconds
Required log reduction:	4.00	Achieved log reduction:	5.00
Test date:	17.06.2013		

Dirty condition

Concentration:	Time:
Required log reduction:	Achieved log reduction:
Test date:	

ROTAVIRUS STRAIN WA - EN 14476

Chemical disinfectants and antiseptics- Viricidal quantitative suspension test for the evaluation chemical disinfectants and antiseptics used in human medicine- Test method and requirements (phase 2, step 1)



Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration:

Time:

Required log reduction:

Achieved log reduction:

Test date:

Dirty condition

Concentration: Undiluted.

Time: 15 seconds

Required log reduction: 4.00

Achieved log reduction: 5.23

Test date: 07.02.2015

VACCINIA VIRUS STRAIN ELSTREE - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 4.00

Achieved log reduction: 5.00

Test date: 07.04.2009

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 4.00

Achieved log reduction: 6.25

Test date: 07.04.2009

Virusesu

BOVINE VIRAL DIARRHEA VIRUS - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration: Undiluted.

Time: 15 seconds

Required log reduction: 4.00

Achieved log reduction: 4.88

Test date: 17.03.2009



Dirty condition

Concentration:	Undiluted.	Time:	15 seconds
Required log reduction:	4.00	Achieved log reduction:	4.51
Test date:	17.03.2009		

CORONAVIRUS - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:		Time:	
Required log reduction:		Achieved log reduction:	
Test date:			

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	4.00
Test date:	25.03.2004		

HEPATITIS B VIRUS - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method:	Suspension test
Test temperature:	20 °C

Clean condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	5.00
Test date:	07.04.2009		

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	6.25
Test date:	07.04.2009		

HEPATITIS C VIRUS - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method:	Suspension test
--------------	-----------------



Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	15 seconds
Required log reduction:	4.00	Achieved log reduction:	4.88
Test date:	17.03.2009		

Dirty condition

Concentration:	Undiluted.	Time:	15 seconds
Required log reduction:	4.00	Achieved log reduction:	4.51
Test date:	17.03.2009		

HERPES SIMPLEX VIRUS TYPE 1 - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	15 seconds
Required log reduction:	4.00	Achieved log reduction:	4.75
Test date:	30.04.2009		

Dirty condition

Concentration:	Undiluted.	Time:	15 seconds
Required log reduction:	4.00	Achieved log reduction:	4.69
Test date:	30.04.2009		

HUMAN IMMUNODEFICIENCY VIRUS - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method: Suspension test
 Test temperature: 20 °C

Clean condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	5.00
Test date:	07.04.2009		

Dirty condition

Concentration:	Undiluted.	Time:	30 seconds
Required log reduction:	4.00	Achieved log reduction:	6.25



Test date: 07.04.2009

INFLUENZA A VIRUS H1N1 - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 4.00

Achieved log reduction: 5.00

Test date: 07.04.2009

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 4.00

Achieved log reduction: 6.25

Test date: 07.04.2009

INFLUENZA A VIRUS H5N1 - RKI / DVV GUIDELINES

Guideline of the German Association for the Control of Viral Diseases (DVV) and of the Robert Koch Institute (RKI) regarding the testing of chemical disinfectants used in human medicine for efficacy against viruses.

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 4.00

Achieved log reduction: 5.00

Test date: 07.04.2009

Dirty condition

Concentration: Undiluted.

Time: 30 seconds

Required log reduction: 4.00

Achieved log reduction: 6.25

Test date: 07.04.2009

INFLUENZA A VIRUS H3N2 - EN 14476

Chemical disinfectants and antiseptics- Viricidal quantitative suspension test for the evaluation chemical disinfectants and antiseptics used in human medicine- Test method and requirements (phase 2, step 1)

Test method: Suspension test

Test temperature: 20 °C

Clean condition

Concentration: Undiluted.

Time: 15 seconds

